

OD-Measurement

- The photometer is activated and set to the desired wavelength.
- 1mL of each sample is transferred to a plastic cuvette, put in the photometer and measured at the correct wavelength.
- 1mL of the solution the microorganism is cultivated in is used as a blank for the measurement.